



TruLab Bicarbonate

In-vitro-Diagnostic for veterinary use only

Assayed quality control material for monitoring assay performance of quantitative in vitro determination of bicarbonate on DiaSys respons[®]910 VET

Order Information

5 9700 99 11 065 3 x 3 mL

Description

TruLab Bicarbonate is a set of three liquid stable controls with one level. The control is based on an aqueous solution.

Storage

The vials of TruLab Bicarbonate both opened and unopened must be stored 35.6 – 77°F.

Stability

Unopened: Until the end of the indicated month of expiry
 Opened: 3 months, if recapped immediately after use

Proper storage and handling of this product must be observed.

Warnings and Precautions

1. Contains sodium azide (0.95 g/L) as preservative. Do not swallow! Avoid contact with skin and mucous membranes.
2. Please refer to the safety data sheets and take the necessary precautions for the use of calibrators and controls.
3. For professional use only!

Preparation

TruLab Bicarbonate is liquid and ready to use.

Procedure

Please refer to the reagent package insert for instructions for use.

Assay Values

TruLab Bicarbonate assay values have been determined by calibration with a reference material in accordance with established protocols. Assay values may vary slightly with different reagents and methodologies used. The assay values listed below are specific for this lot number of control only.

Waste Management

Please refer to local legal requirements.

Manufacturer

DiaSys Diagnostic Systems GmbH
 Alte Strasse 9
 65558 Holzheim Germany

	Lot No.	Expiry Date	Assay Values	Range
TruLab Bicarbonate	33619	2023-11	20.0 mmol/L	16.0 – 24.0 mmol/L