




TruLab L Level 1

LOT 33893



In-vitro-diagnostic for veterinary use only



2025-09

Cat. No. 5 9020 99 11 065 (3 x 3 mL)

Value sheet version: 1

Short Name / Analyte	Method	Reagent ID	Value	Range	Unit			
APOA1 Apolipoprotein A1	Immunoturbidimetric	1 7102	126 1.26	101 - 151 1.01 - 1.51	mg/dL g/L			
APOB Apolipoprotein B	Immunoturbidimetric	1 7112	88.6 0.886	70.9 - 106 0.709 - 1.06	mg/dL g/L			
CHOL Cholesterol	CHOD-PAP	1 1300 +	176	151 - 200	mg/dL			
		1 1350	4.55 1.76	3.91 - 5.18 1.51 - 2.00	mmol/L g/L			
HDLC HDL-Cholesterol	Immunoinhibition	1 3521	48.5 1.25 0.485	38.8 - 58.2 1.00 - 1.50 0.388 - 0.582	mg/dL mmol/L g/L			
			Direct, colorimetric	1 3561	46.0 1.19 0.460	36.8 - 55.2 0.952 - 1.43 0.368 - 0.552	mg/dL mmol/L g/L	
					CHOD-PAP Precipitant	1 1300 +	43.5	34.8 - 52.1
	1 3540	1.12 0.435				0.899 - 1.35 0.348 - 0.521	mmol/L g/L	
	LDLC LDL-Cholesterol	Homogenous Select, TruCal Lipid	1 4121	100 2.59 1.00	80.2 - 120 2.08 - 3.11 0.802 - 1.20	mg/dL mmol/L g/L		
				Direct, colorimetric	1 4131	103 2.66 1.03	82.2 - 123 2.12 - 3.19 0.822 - 1.23	mg/dL mmol/L g/L
CHOD-PAP Precipitant						1 1300 +	117	93.6 - 140
			1 4330	3.02 1.17	2.42 - 3.63 0.936 - 1.40	mmol/L g/L		
Lp-PLA₂ Lipoprotein-associated phospholipase A2		Colorimetric	1 7181	335 5.58	268 - 402 4.47 - 6.70	U/L µkat/L		
	NEFA Non-esterified fatty acids			Enzymatic	1 5781	16.2 0.573 0.162	12.9 - 19.4 0.458 - 0.687 0.129 - 0.194	mg/dL mmol/L g/L
PL Phospholipids		Enzymatic	1 5741			212 2.73 2.12	169 - 254 2.19 - 3.28 1.69 - 2.54	mg/dL mmol/L g/L
						TRIG Triglycerides	GPO-PAP	1 5760
	1 5710			1.13 1.00	0.902 - 1.35 0.801 - 1.20			




DiaSys Diagnostic Systems GmbH
 Alte Strasse 9 65558 Holzheim Germany
www.diasys-diagnostics.com



TruLab L Level 1

LOT 33893



In-vitro-diagnostic for veterinary use only



2025-09

Cat. No. 5 9020 99 11 065 (3 x 3 mL)

respons[®] 910_{vet}

Value sheet version: 1

Short Name / Analyte	Method	Reagent ID	Value	Range	Unit	
HDLC HDL-Cholesterol	Immunoinhibition	1 3521	48.5	38.8 - 58.2	mg/dL	
			1.25	1.00 - 1.50	mmol/L	
			0.485	0.388 - 0.582	g/L	
	Direct, colorimetric	1 3561	46.0	36.8 - 55.2	mg/dL	
			1.19	0.952 - 1.43	mmol/L	
			0.460	0.368 - 0.552	g/L	
LDLC LDL-Cholesterol	Homogenous Select, TruCal Lipid	1 4121	100	80.2 - 120	mg/dL	
			2.59	2.08 - 3.11	mmol/L	
			1.00	0.802 - 1.20	g/L	
	Direct, colorimetric	1 4131	103	82.2 - 123	mg/dL	
			2.66	2.12 - 3.19	mmol/L	
			1.03	0.822 - 1.23	g/L	
NEFA Non-esterified fatty acids	Enzymatic	1 5781	16.2	12.9 - 19.4	mg/dL	
			0.573	0.458 - 0.687	mmol/L	
			0.162	0.129 - 0.194	g/L	
TRIG Triglycerides	GPO-PAP	1 5760	100	80.1 - 120	mg/dL	
			1 5710	1.13	0.902 - 1.35	mmol/L
				1.00	0.801 - 1.20	g/L



DiaSys Diagnostic Systems GmbH
Alte Strasse 9 65558 Holzheim Germany
www.diasys-diagnostics.com



TruLab L Level 1

LOT 33893



In-vitro-diagnostic for veterinary use only



2025-09

Cat. No. 5 9020 99 11 065 (3 x 3 mL)

respons[®]920_{vet}

Value sheet version: 1

Short Name / Analyte	Method	Reagent ID	Value	Range	1SD	Unit	
HDLC HDL-Cholesterol	Immunoinhibition	1 3521	48.5	38.8 - 58.2	3.23	mg/dL	
			1.25	1.00 - 1.50	0.084	mmol/L	
			0.485	0.388 - 0.582	0.032	g/L	
	Direct, colorimetric	1 3561	46.0	36.8 - 55.2	3.07	mg/dL	
			1.19	0.952 - 1.43	0.079	mmol/L	
			0.460	0.368 - 0.552	0.031	g/L	
LDLC LDL-Cholesterol	Homogenous Select, TruCal Lipid	1 4121	98.2	78.6 - 118	6.55	mg/dL	
			2.54	2.03 - 3.05	0.169	mmol/L	
			0.982	0.786 - 1.18	0.065	g/L	
	Direct, colorimetric	1 4131	104	83.4 - 125	6.95	mg/dL	
			2.70	2.16 - 3.24	0.180	mmol/L	
			1.04	0.834 - 1.25	0.070	g/L	
TRIG Triglycerides	GPO-PAP	1 5760	100	80.1 - 120	6.67	mg/dL	
			1 5710	1.13	0.902 - 1.35	0.075	mmol/L
				1.00	0.801 - 1.20	0.067	g/L



DiaSys Diagnostic Systems GmbH
Alte Strasse 9 65558 Holzheim Germany
www.diasys-diagnostics.com



TruLab L Level 1

LOT 33893



In-vitro-diagnostic for veterinary use only



2025-09

Cat. No. 5 9020 99 11 065 (3 x 3 mL)

respons[®] 940_{vet}

Value sheet version: 1

Short Name / Analyte	Method	Reagent ID	Value	Range	1SD	Unit	
HDLC HDL-Cholesterol	Immunoinhibition	1 3521	48.5	38.8 - 58.2	3.23	mg/dL	
			1.25	1.00 - 1.50	0.084	mmol/L	
			0.485	0.388 - 0.582	0.032	g/L	
	Direct, colorimetric	1 3561	46.0	36.8 - 55.2	3.07	mg/dL	
			1.19	0.952 - 1.43	0.079	mmol/L	
			0.460	0.368 - 0.552	0.031	g/L	
LDLC LDL-Cholesterol	Homogenous Select, TruCal Lipid	1 4121	98.2	78.6 - 118	6.55	mg/dL	
			2.54	2.03 - 3.05	0.169	mmol/L	
			0.982	0.786 - 1.18	0.065	g/L	
	Direct, colorimetric	1 4131	104	83.4 - 125	6.95	mg/dL	
			2.70	2.16 - 3.24	0.180	mmol/L	
			1.04	0.834 - 1.25	0.070	g/L	
TRIG Triglycerides	GPO-PAP	1 5760	100	80.1 - 120	6.67	mg/dL	
			1 5710	1.13	0.902 - 1.35	0.075	mmol/L
				1.00	0.801 - 1.20	0.067	g/L



DiaSys Diagnostic Systems GmbH
Alte Strasse 9 65558 Holzheim Germany
www.diasys-diagnostics.com