




# TruLab L Level 2

**LOT** 30762



In-vitro-diagnostic for veterinary use only



2024-02

Cat. No. 5 9030 99 11 065 ( 3 x 3 mL)

Value sheet version: 5

Short Name / Analyte	Method	Reagent ID	Value	Range	Unit
<b>APOA1</b> Apolipoprotein A1	Immunoturbidimetric	1 7102	203	162 - 244	mg/dL
			2.03	1.62 - 2.44	g/L
<b>APOB</b> Apolipoprotein B	Immunoturbidimetric	1 7112	157	125 - 188	mg/dL
			1.57	1.25 - 1.88	g/L
<b>CHOL</b> Cholesterol	CHOD-PAP	1 1300 +	289	248 - 329	mg/dL
		1 1350	7.47	6.42 - 8.51	mmol/L
			2.89	2.48 - 3.29	g/L
<b>HDLC</b> HDL-Cholesterol	Immunoinhibition	1 3521	70.9	56.7 - 85.1	mg/dL
			1.83	1.47 - 2.20	mmol/L
			0.709	0.567 - 0.851	g/L
	Direct, colorimetric	1 3561	65.2	52.2 - 78.2	mg/dL
			1.69	1.35 - 2.02	mmol/L
			0.652	0.522 - 0.782	g/L
CHOD-PAP Precipitant	1 1300 +	68.2	54.5 - 81.8	mg/dL	
		1 3540	1.76	1.41 - 2.12	mmol/L
			0.682	0.545 - 0.818	g/L
<b>LDLC</b> LDL-Cholesterol	Homogenous Select, TruCal Lipid	1 4121	158	127 - 190	mg/dL
			4.09	3.27 - 4.91	mmol/L
			1.58	1.27 - 1.90	g/L
	Direct, colorimetric	1 4131	173	138 - 208	mg/dL
			4.47	3.58 - 5.37	mmol/L
			1.73	1.38 - 2.08	g/L
CHOD-PAP Precipitant	1 1300 +	190	152 - 228	mg/dL	
		1 4330	4.92	3.93 - 5.90	mmol/L
			1.90	1.52 - 2.28	g/L
<b>Lp-PLA<sub>2</sub></b> Lipoprotein-associated phospholipase A2	Colorimetric reconstitution volume.: 1.0 mL	1 7181	1076	861 - 1291	U/L
			17.9	14.3 - 21.5	µkat/L
			810	648 - 972	U/L
	Colorimetric reconstitution volume.: 1.5 mL	1 7181	13.5	10.8 - 16.2	µkat/L
			53.8	43.0 - 64.5	mg/dL
			1.90	1.52 - 2.28	mmol/L
<b>NEFA</b> Non-esterified fatty acids	Enzymatic	1 5781	0.538	0.430 - 0.645	g/L
<b>PL</b> Phospholipids	Enzymatic	1 5741	313	250 - 375	mg/dL
			4.03	3.23 - 4.84	mmol/L
			3.13	2.50 - 3.75	g/L
<b>TRIG</b> Triglycerides	GPO-PAP	1 5760	265	212 - 318	mg/dL
		1 5710	2.99	2.39 - 3.59	mmol/L
			2.65	2.12 - 3.18	g/L




DiaSys Diagnostic Systems GmbH  
Alte Strasse 9 65558 Holzheim Germany  
www.diasys-diagnostics.com



# TruLab L Level 2

**LOT** 30762



In-vitro-diagnostic for veterinary use only



2024-02

Cat. No. 5 9030 99 11 065 ( 3 x 3 mL)

**respons<sup>®</sup> 910<sub>Net</sub>**

Value sheet version: 5

Short Name / Analyte	Method	Reagent ID	Value	Range	Unit	
<b>APOA1</b> Apolipoprotein A1	Immunoturbidimetric	1 7102	203 2.03	162 - 244 1.62 - 2.44	mg/dL g/L	
<b>APOB</b> Apolipoprotein B	Immunoturbidimetric	1 7112	157 1.57	125 - 188 1.25 - 1.88	mg/dL g/L	
<b>CHOL</b> Cholesterol	CHOD-PAP	1 1300 +	289	248 - 329	mg/dL	
		1 1350	7.47 2.89	6.42 - 8.51 2.48 - 3.29	mmol/L g/L	
<b>HDLC</b> HDL-Cholesterol	Immunoinhibition	1 3521	70.9	56.7 - 85.1	mg/dL	
			1.83 0.709	1.47 - 2.20 0.567 - 0.851	mmol/L g/L	
<b>LDLC</b> LDL-Cholesterol	Direct, colorimetric	1 3561	65.2	52.2 - 78.2	mg/dL	
			1.69	1.35 - 2.02	mmol/L	
			0.652	0.522 - 0.782	g/L	
<b>LDLC</b> LDL-Cholesterol	Homogenous Select, TruCal Lipid	1 4121	158	127 - 190	mg/dL	
			4.09	3.27 - 4.91	mmol/L	
			1.58	1.27 - 1.90	g/L	
<b>Lp-PLA<sub>2</sub></b> Lipoprotein-associated phospholipase A2	Colorimetric reconstitution volume.: 1.0 mL	1 7181	1076	861 - 1291	U/L	
			17.9	14.3 - 21.5	µkat/L	
<b>NEFA</b> Non-esterified fatty acids	Enzymatic	1 5781	53.8	43.0 - 64.5	mg/dL	
			1.90	1.52 - 2.28	mmol/L	
			0.538	0.430 - 0.645	g/L	
<b>PL</b> Phospholipids	Enzymatic	1 5741	313	250 - 375	mg/dL	
			4.03	3.23 - 4.84	mmol/L	
			3.13	2.50 - 3.75	g/L	
<b>TRIG</b> Triglycerides	GPO-PAP	1 5760 1 5710	265	212 - 318	mg/dL	
			2.99	2.39 - 3.59	mmol/L	
			2.65	2.12 - 3.18	g/L	



DiaSys Diagnostic Systems GmbH  
Alte Strasse 9 65558 Holzheim Germany  
www.diasys-diagnostics.com



# TruLab L Level 2

**LOT** 30762



In-vitro-diagnostic for veterinary use only



2024-02

Cat. No. 5 9030 99 11 065 ( 3 x 3 mL)

**respons<sup>®</sup>920<sub>Net</sub>**

Value sheet version: 5

Short Name / Analyte	Method	Reagent ID	Value	Range	1SD	Unit
<b>APOA1</b> Apolipoprotein A1	Immunoturbidimetric	1 7102	203 2.03	162 - 244 1.62 - 2.44	13.5 0.135	mg/dL g/L
<b>APOB</b> Apolipoprotein B	Immunoturbidimetric	1 7112	157 1.57	125 - 188 1.25 - 1.88	10.4 0.104	mg/dL g/L
<b>CHOL</b> Cholesterol	CHOD-PAP	1 1300 +	289	248 - 329	13.5	mg/dL
		1 1350	7.47 2.89	6.42 - 8.51 2.48 - 3.29	0.349 0.135	mmol/L g/L
<b>HDLC</b> HDL-Cholesterol	Immunoinhibition	1 3521	70.9	56.7 - 85.1	4.73	mg/dL
			1.83	1.47 - 2.20	0.122	mmol/L
			0.709	0.567 - 0.851	0.047	g/L
	Direct, colorimetric	1 3561	65.2 1.69 0.652	52.2 - 78.2 1.35 - 2.02 0.522 - 0.782	4.35 0.112 0.043	mg/dL mmol/L g/L
<b>LDLC</b> LDL-Cholesterol	Homogenous Select, TruCal Lipid	1 4121	151	121 - 181	10.0	mg/dL
			3.90	3.12 - 4.68	0.260	mmol/L
			1.51	1.21 - 1.81	0.100	g/L
	Direct, colorimetric	1 4131	173	138 - 208	11.5	mg/dL
			4.47 1.73	3.58 - 5.37 1.38 - 2.08	0.298 0.115	mmol/L g/L
<b>TRIG</b> Triglycerides	GPO-PAP	1 5760	265	212 - 318	17.7	mg/dL
		1 5710	2.99	2.39 - 3.59	0.199	mmol/L
			2.65	2.12 - 3.18	0.177	g/L



DiaSys Diagnostic Systems GmbH  
Alte Strasse 9 65558 Holzheim Germany  
www.diasys-diagnostics.com



# TruLab L Level 2

**LOT** 30762



Cat. No. 5 9030 99 11 065 ( 3 x 3 mL)



2024-02

In-vitro-diagnostic for veterinary use only

**respons<sup>®</sup> 940<sub>vet</sub>**

Value sheet version: 5

Short Name / Analyte	Method	Reagent ID	Value	Range	1SD	Unit
<b>APOA1</b> Apolipoprotein A1	Immunoturbidimetric	1 7102	203 2.03	162 - 244 1.62 - 2.44	13.5 0.135	mg/dL g/L
<b>APOB</b> Apolipoprotein B	Immunoturbidimetric	1 7112	157 1.57	125 - 188 1.25 - 1.88	10.4 0.104	mg/dL g/L
<b>CHOL</b> Cholesterol	CHOD-PAP	1 1300 +	289	248 - 329	13.5	mg/dL
		1 1350	7.47 2.89	6.42 - 8.51 2.48 - 3.29	0.349 0.135	mmol/L g/L
<b>HDLC</b> HDL-Cholesterol	Immunoinhibition	1 3521	70.9	56.7 - 85.1	4.73	mg/dL
			1.83	1.47 - 2.20	0.122	mmol/L
			0.709	0.567 - 0.851	0.047	g/L
	Direct, colorimetric	1 3561	65.2 1.69 0.652	52.2 - 78.2 1.35 - 2.02 0.522 - 0.782	4.35 0.112 0.043	mg/dL mmol/L g/L
<b>LDLC</b> LDL-Cholesterol	Homogenous Select, TruCal Lipid	1 4121	151	121 - 181	10.0	mg/dL
			3.90	3.12 - 4.68	0.260	mmol/L
			1.51	1.21 - 1.81	0.100	g/L
	Direct, colorimetric	1 4131	173	138 - 208	11.5	mg/dL
			4.47 1.73	3.58 - 5.37 1.38 - 2.08	0.298 0.115	mmol/L g/L
<b>TRIG</b> Triglycerides	GPO-PAP	1 5760	265	212 - 318	17.7	mg/dL
		1 5710	2.99	2.39 - 3.59	0.199	mmol/L
			2.65	2.12 - 3.18	0.177	g/L



DiaSys Diagnostic Systems GmbH  
Alte Strasse 9 65558 Holzheim Germany  
www.diasys-diagnostics.com